Comprehensive CT Evaluation of Valve & Vessels

Practical Issues and Clinical Outcomes

Won-Jang Kim, MD, PhD

Clinical Assistant Professor of Medicine, Heart Institute, Asan Medical Center, Seoul, Korea

Structural and Valvular Intervention, Sejong General Hospital, Korea






Evaluation of Access Routes

Reduce Vascular Injury



Femoral Artery Puncture under Fluoroscopic Guidance




Initial Heofemoral Aortography



Baseline Angiography & CT






Made by Adw 4.5, GE healthcare system

Difficulty in Advancement Severe calcific small vessel








lleofemoral Artery Evaluation





lleofemoral Artery Evaluation





CT Screening Can Help Reduce Vascular Injury Rates

Variables	2009	2010	P value
MDCT Screening	44%	69%	<0.01
Ultrasound-guided puncture	0	37%	<0.01
Sheath size >19F	40%	2%	<0.01
Expandable sheath	12%	18%	0.33
MLD < external sheath diameter	77%	30%	<0.01
All vascular complications	32%	9%	<0.01

Criteria to Predict Vascular Complications

	SFAR	
≥1.05 (n=55)	<1.05 (n=72)	P Value
41.8%	16.7%	<0.001
30.9%	6.9%	0.001
10.9%	9.7%	0.827
27.3%	12.5%	0.035
20.0%	2.8%	0.002
20.0%	6.9%	0.033
18.2%	4.2%	0.016
	41.8% 30.9% 10.9% 27.3% 20.0%	≥1.05 (n=55) <1.05 (n=72) 41.8% 16.7% 30.9% 6.9% 10.9% 9.7% 27.3% 12.5% 20.0% 2.8% 6.9%

Decrease Complications with Experiences and Device Developments

Edwards Cases	RF I or III N=9	NovaFlex N=21	
Procedural success	8 (88.9%)	21 (100%)	
Mortality	0	0	
Stroke	0	1 (4.8%)	
Permanent pacemaker	0	0	
Vascular complication			
Access site	1 (11.1%)	0	
Iliac artery perforation	1 (11.1%)	0	
Device embolization	2 (22.2%)	1 (4.8%)	

Annulus sizing


Cannot be emphasized enough...

For successful procedure & reduce complications







PPM or Rupture vs. PVL






Aortic Annulus on CT
























Circular Annulus is Very Small Proportion


Distribution of D_{max}/D_{min} from 164 TAVI patients Courtesy of Dr. Piazza and Prof. Lange, German Heart Center, Munich Germany

A Limitation of 2-D Image




It is possible a true diameter is not measured due to the imaging plane acquired


Aortic Annulus on MSCT



Coronal Image



Basal Ring




Oblique Sagittal Image





New CT Parameters





Reliability Comparison


TEE vs. CT Variables (N=30, Preliminary AMC Data)

TCC


CT measurements for annulus are usually larger than TEE measurements. CT perimeter & area measurements are most reproducible.

by ICC (1)						
(2)	0.51 (0.40-0.62)	0.93 (0.84-0.97)	0.95 (0.88-0.97)	0.96 (0.89-0.99)	0.93 (0.83-0.96)	0.95 (0.86-0.98)


CT Measures Can Predict PVL




- Valve stent diameter –
 Mean annular diameter_{CT} =
 AUC 0.84
- Valve stent diameter –
 Area-derived annular
 diameter_{CT} = AUC 0.86
- Valve stent area/ Annular area_{CT} = AUC 0.87



CT Annular Measures Appear more Predictive than 3-D Echo for PVL




- CT Dmean Annulus AUC= 0.82
- 3D TEE Mean Annulus AUC = 0.68
- 2D TEE AUC = 0.52



Stretching Index vs. PPM 100 80 80 Sensitivity 86.96% Sensitivity Sensitivity (%) Specificity 94.68% 60 Specificity 60 **PPV 80%** NPV 96.74% 40 40 Accuracy 93.2% Cut-off =1.13 20 20 AUC=0.91 95% CI=0.820-0.992 40 60 80 100 1.2 1.27 1.34 1.0 1.06 1.13 -1.08 100-Specificity Stretching Index using Perimeter **Depth vs. PPM** 100 100 F Sensitivity 60.87% 80 80 Specificity 74.47% Sensitivity (%) PPV 35.14% Sensitivity 60 60 NPV 87.5% Specificity Accuracy 70.94% 40 40 Cut-off = 7.8 mm20 20 AUC=0.70 95% CI=0.593-0.806 0 20 40 60 80 100 Asian Registry, in submitting NCC to Basal Skirt of Device (mm) 100-Specificity

Best Combination for Prevention of Permanent Pacemaker



1.13
Stretching Index


Predictors of aortic root rupture

	Univariate		Multivariable	
	Odds Ratio (95%CI)	P value	Odds Ratio (95%CI)	P value
LVOT calcifications moderate/severe	6.03 (2.35-15.45)	<0.001	12.45 (2.97-52.15)	0.001
Prosthesis oversizing ≥ 20%	8.76 (3.19-24.09)	<0.001	23.17 (4.77- 45.71)	<0.001
Balloon post-dilation	9.00 (2.59-22.08)	0.001	10.40 (1.54-30.46)	0.016

Adjusted for gender, MDCT annular area, MDCT LVOT area, presence of MDCT LVOT moderate to severe calcification, presence of MDCT aortic valve moderate to severe calcification, presence of prosthesis oversizing ≥20%, MDCT SV maximal diameter, and balloon post-dilation.

Annular Sizing for TAVR

Measurement of Annulus Dimensions





CT Sizing for CoreValve

Valve Size	Diameter	Perimeter	Cover Index
31mm	29mm	91.1	6.45%
31mm	28mm	88	10.30%
31mm	27 mm	84.8	12.90%
31mm	26mm	81.7	16.13%
29mm	27mm	84.8	6.90%
29mm	26mm	81.7	10.30%
29mm	25mm	78.5	13.80%
29mm	24mm	75.4	17.20%
26mm	23mm	72.3	11.50%
26mm	22mm	69.1	15.40%
26mm	21mm	66	19.20%
26mm	20mm	62.8	23.10%


CT Sizing for Edwards Valve


Annular Area (mm²)	Edwards valve size (mm)
230 - 300	20
310 - 320	20 or 23
330 - 400	23
410	23 or 26
420 - 510	26
520	26 or 29
530 - 660	29

Impact of Underfilling on Expansion In Vitro

	Balloon	1 ml	2 ml	3 ml	4 ml
	volume	underfilled	underfilled	underfilled	underfilled
Novaflex					
20-mm THV	11 ml	-9.1%	-18.2%*	-27.3%*	-36.4%*
23-mm THV	17 ml	-5.9%	-11.8%	-17.6%*	-23.5%*
26-mm THV	22 ml	-4.5%	-9.0%	-13.6%	-18.2%*
29-mm THV	33 ml	-3.0%	-6.1%	-9.1%	-12.1%
Ascendra					
23-mm THV	16 ml	-6.3%	-12.5%	-18.8%*	-25.0%*
26-mm THV	20 ml	-5.0%	-10.0%	-15.0%	-20.0%*
29-mm THV	30 ml	-3.3%	-6.7%	-10.0%	-13.3%

Aortic Root Anatomy and Distances






	Width	Height	For annulus diameter	Height of skirt
Edward SAPIEN XT TM	23mm	14.3mm	18-22mm	10.1/7.74mm
	26mm	17.2mm	21-25mm	11.4/8.67mm
CoreValve Revalving™	26mm	53mm	20-23mm	12mm
	29mm	55mm	23-27mm	12mm




Left main height




 Curved MPR or max oblique coronal view

DIASTOLIC PHASE

SYSTOLIC PHASE





Potential Mechanisms of Coronary Ostial Obstruction

- Impingement of ostia by THV support structure
- High positioning of sealing cuff
- Embolization of atheroma, calcium, thrombus, air or vegetation
- Oversizing of THV
- Dissection of aortic root
- Displacement of native aortic leaflets towards coronary ostia

Recommended annulus to ostial height: > 10 mm for Sapien 23 and > 11 mm for Sapien 26

Coronary Height





Bicuspid AV




Right Coronary Artery





Navigator For Transapical Approach



Direction of Puncture or Wire

Aortic Valve Morphology & Amount of Calcium

Scanty calcium

Heavy eccentric calcium

Echocardiographic findings

Calcificated structure is enemy of Echo




TEE

TTE





Amount of Cuspid Calcification



Scanty of Calcium

Heavy Eccentric Calcium






Heavy Eccentric Calcium





Center

Heavy Eccentric Calcium






23 mm Edward Valve






Heavy Eccentric Calcium: Extent












Valve positioning



Aortic Valve Plane by CT Scan













CT vs 3-D Angio CT for Angle Prediction




Valve Placement






Assisting with Valve Positioning



DynaCT Image Acquisition with rapid pacing






Courtesy Siemens Systems





Merged Imaging Tools



Valve deployment under DynaCT







CoreValve

Courtesy by Alois Nöttling Siemens
Courtesy by Brockmann German Heart Center Munich






Follow up evaluation




Examples of ConformabilityCoreValve Cases






Follow Up Image







No Valve Migration, Fracture, Circumferentiality



Major Roles of CT in TAVI

- Ileofemoral Arterial Sytem : Size, Calcification, Tortuosity, Plaques
- Annulus size measurement
- 3D annular & root morphology & dimensions
- Relationship of annulus to both coronary ostia
- Amounts of calcium in valve
- Valve positioning during implantation
- Post TAVI assessment

